



CERTIFICATE OF ANALYSIS

Sodium Eluant for Amino Acid Analysis

Na740

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na740

Lot Number: 007017

Test Start Date: 16 Jul 2010

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 7.55 ± 0.06 @ 25 ± 0.4 °C

2. Retention time of L-histidine meets specifications (Y/N).

3. Resolution of Tyr/Phe ≥ 75 % (Y/N).


Actual Values

7.575

Y

Y

Calibration Standard: Collagen Hydrolysate, PN 012506C, LN 002006

Analysis Certified by:  **Date:** 21 Jul 2010

VP Ops

Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE