



# CERTIFICATE OF ANALYSIS

## Sodium Eluant for Amino Acid Analysis

Na740

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na740

Lot Number: 008016

Test Start Date: 13 AUG 2010

Primary Application: Sodium-based amino acid analysis.

**Acceptance Criteria:**

1. pH =  $7.55 \pm 0.06$  @  $25 \pm 0.4$  °C

2. Retention time of L-histidine meets specifications (Y/N).

3. Resolution of Tyr/Phe  $\geq 75$  % (Y/N).

**Actual Values**

7.565

Y

Y

Calibration Standard: Collagen Hydrolysate, PN 012506C, LN 002006

Analysis Certified by:  Date: 23 Aug 2010

VP OPR

Title

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE