



CERTIFICATE OF ANALYSIS

Sodium Eluant for Amino Acid Analysis

Na740

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na740

Lot Number: 010012

Test Start Date: 22 Oct 2010

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH = 7.55 ± 0.06 @ 25 ± 0.4 °C

Actual Values

7.562

2. Retention time of L-histidine meets specifications (Y/N).

Y

3. Resolution of Tyr/Phe ≥ 75 % (Y/N).

Y

Calibration Standard: Collagen Hydrolysate, PN 012506C, LN 002006

Analysis Certified by:  Date: 27 Oct 2010

VP Opr

Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE