



# CERTIFICATE OF ANALYSIS

## Sodium Eluant for Amino Acid Analysis

Na740

**Product:** Sodium Eluant for Amino Acid Analysis

**Part Number:** Na740

**Lot Number:** 103034

**Production Date:** 23 MAR 2011

**Expiration Date:** 22 MAR 2013

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

1. pH > 7.40 @ 25 ± 0.4 °C (Y/N)

Y

2. Retention time of L-histidine meets specifications (Y/N).

Y

3. Resolution of Tyr/Phe ≥ 75 % (Y/N).

Y

**Calibration Standard:** Collagen Hydrolysate, PN 012506C,

**Analysis Certified by:**

*[Signature]*  
VP Oper

Title

**Date:**

13 Apr 2011

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE