



CERTIFICATE OF ANALYSIS

Sodium Eluant for Amino Acid Analysis

Na740

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na740

Lot Number: 104014

Production Date: 12 APR 2011

Expiration Date: 11 APR 2013

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH > 7.40 @ 25 ± 0.4 °C (Y/N)

Y

2. Retention time of L-histidine meets specifications (Y/N).

Y

3. Resolution of Tyr/Phe ≥ 75 % (Y/N).

Y

Calibration Standard: Collagen Hydrolysate, PN 012506C,

Analysis Certified by: _____

John Murphy
VP Ops

Title

Date: 18 Apr 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE