



CERTIFICATE OF ANALYSIS

Sodium Eluant for Amino Acid Analysis

Na740

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na740

Lot Number: 104028

Production Date: 22 APR 2011

Expiration Date: 21 APR 2013

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH > 7.40 @ 25 ± 0.4 °C (Y/N) Y
2. Retention time of L-histidine meets specifications (Y/N). Y
3. Resolution of Tyr/Phe ≥ 75 % (Y/N). Y

Calibration Standard: Collagen Hydrolysate, PN 012506C,

Analysis Certified by: *Jim Murphy* Date: 25 Apr 2011

VP Opr
Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE