



**CERTIFICATE OF ANALYSIS**  
**Sodium Eluant for Amino Acid Analysis**  
**Na740**

**Product:** Sodium Eluant for Amino Acid Analysis  
**Part Number:** Na740  
**Lot Number:** 106008  
**Production Date:** 09 Jun 2011  
**Expiration Date:** 08 Jun 2013

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

- 1. pH > 7.40 @ 25 ± 0.4 °C (Y/N) Y
- 2. Retention time of L-histidine meets specifications (Y/N). Y
- 3. Resolution of Tyr/Phe ≥ 75 % (Y/N). Y

**Calibration Standard:** Collagen Hydrolysate, PN 012506C,

**Analysis Certified by:** *John Murphy* **Date:** 21 Jun 2011  
VP OR  
Title

**ACCEPTED**  
PICKERING LABORATORIES, INC  
QUALITY ASSURANCE

Rev. 9, 04/11/2011