



CERTIFICATE OF ANALYSIS

Sodium Eluant for Amino Acid Analysis

Na740

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na740

Lot Number: 108010

Production Date: 18 AUG 2011

Expiration Date: 17 AUG 2013

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH > 7.40 @ 25 ± 0.4 °C (Y/N) /

2. Retention time of L-histidine meets specifications (Y/N). /

3. Resolution of Tyr/Phe ≥ 75 % (Y/N). /

Calibration Standard: Collagen Hydrolysate, PN 012506C,

Analysis Certified by: _____

Steve Murphy
VP OPR
Title

Date: 20 Sep 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 9, 04/11/2011