



**CERTIFICATE OF ANALYSIS**  
**Sodium Eluant for Amino Acid Analysis**  
**Na740**


**Product:** Sodium Eluant for Amino Acid Analysis  
**Part Number:** Na740  
**Lot Number:** 109036  
**Production Date:** 27 Sep 2011  
**Expiration Date:** 26 Sep 2013

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

- |  |              |
|--|--------------|
| 1. pH > 7.40 @ 25 ± 0.4 °C (Y/N)                             | <u>  4  </u> |
| 2. Retention time of L-histidine meets specifications (Y/N). | <u>  4  </u> |
| 3. Resolution of Tyr/Phe meets specification (Y/N).          | <u>  4  </u> |
| 4. No peaks in baseline (Y/N)                                | <u>  4  </u> |

**Calibration Standard:** Collagen Hydrolysate, PN 012506C,

**Analysis Certified by:**  **Date:** 05 Oct 2011  
VP OPR  
Title

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 9, 04/11/2011