



## CERTIFICATE OF ANALYSIS

### Sodium Eluant for Amino Acid Analysis

Na740

**Product:** Sodium Eluant for Amino Acid Analysis

**Part Number:** Na740

**Lot Number:** 111013

**Production Date:** 09 Nov 2011


**Expiration Date:** 09 Nov 2013

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

1. pH > 7.40 @ 25 ± 0.4 °C (Y/N) Y
2. Retention time of L-histidine meets specifications (Y/N). Y
3. Resolution of Tyr/Phe meets specification (Y/N). Y
4. No peaks in baseline (Y/N) Y

**Calibration Standard:** Collagen Hydrolysate, PN 012506C,

**Analysis Certified by:**  **Date:** 21 Nov 2011  
VP Oper  
Title

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 9, 04/11/2011