



CERTIFICATE OF ANALYSIS

Sodium Eluant for Amino Acid Analysis

Na740

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na740

Lot Number: 112014

Production Date: 15 Dec 2011

Expiration Date: 14 Dec 2013

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH > 7.40 @ 25 ± 0.4 °C (Y/N)
2. Resolution meets specifications:
 - Tyrosine/Phenylalanine (Y/N)
3. Retention time of L-Histidine meets specifications (Y/N)
4. Baseline: no peaks (Y/N)

Y
Y
Y
Y

Calibration Standard: Collagen Hydrolysate, PN 012506C,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

Jim Young
VP OPER
Title

Date:

05 MAR 2012

Rev. 10, 11/30/2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE