



## CERTIFICATE OF ANALYSIS

### Sodium Eluant for Amino Acid Analysis

**Na740**

**Product:** Sodium Eluant for Amino Acid Analysis

**Part Number:** Na740

**Lot Number:** 201029

**Production Date:** 20 JAN 2012

**Expiration Date:** 19 JAN 2014

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

1. pH > 7.40 @ 25 ± 0.4 °C (Y/N)
2. Resolution meets specifications:
  - Tyrosine/Phenylalanine (Y/N)
3. Retention time of L-Histidine meets specifications (Y/N)
4. Baseline: no peaks (Y/N)

Y  
Y  
Y  
Y

**Calibration Standard:** Collagen Hydrolysate, PN 012506C,

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** *Jim Murphy*  
VP OPER  
Title

**Date:** 05 MAR 2012

Rev. 10, 11/30/2011

**ACCEPTED**  
PICKERING LABORATORIES, INC  
QUALITY ASSURANCE