



CERTIFICATE OF ANALYSIS

Sodium Eluant for Amino Acid Analysis

Na740

Product: Sodium Eluant for Amino Acid Analysis

Part Number: Na740

Lot Number: 2510049

Production Date: 23 Oct 2025

Expiration Date: 23 Oct 2027

Primary Application: Sodium-based amino acid analysis.

Acceptance Criteria:

1. pH > 7.40 @ 25 ± 0.4 °C (Y/N)	<u>Y</u>
2. Resolution meets specifications:	
• Tyrosine/Phenylalanine (Y/N)	<u>Y</u>
3. Retention time of L-Histidine meets specifications (Y/N)	<u>Y</u>
4. Baseline: no peaks (Y/N)	<u>Y</u>

Calibration Standard: Collagen Hydrolysate, **PN** 012506C,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:	<u>David Mazawa</u>	Date:	<u>29 Oct 2025</u>
Title:	<u>QC Manager</u>		