



## CERTIFICATE OF ANALYSIS

### Sodium Eluant for Amino Acid Analysis

**Na740**

**Product:** Sodium Eluant for Amino Acid Analysis

**Part Number:** Na740

**Lot Number:** 304021

**Production Date:** 22 April 2013

**Expiration Date:** 22 April 2015

**Primary Application:** Sodium-based amino acid analysis.

**Acceptance Criteria:**

- |   |          |
|---|----------|
| 1. pH > 7.40 @ 25 ± 0.4 °C (Y/N)                            | <u>y</u> |
| 2. Resolution meets specifications:                         |          |
| • Tyrosine/Phenylalanine (Y/N)                              | <u>y</u> |
| 3. Retention time of L-Histidine meets specifications (Y/N) | <u>y</u> |
| 4. Baseline: no peaks (Y/N)                                 | <u>y</u> |

**Calibration Standard:** Collagen Hydrolysate, **PN** 012506C,

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** 29 May 2013  
President  
**Title**