



# CERTIFICATE OF ANALYSIS

## O-Phthalaldehyde (OPA)

**O120**

**Product:** o-phthalaldehyde (OPA)

**Part Number:** O120

**Lot Number:** 2404033

**Production Date:** 16 April 2024

**Expiration Date:** 16 April 2027

**Primary Application:** Reagent for Post-column analysis.

**Acceptance Criteria:**

- |   |          |
|---|----------|
| 1. Color : Yellow (Y/N))                            | <u>Y</u> |
| 2. Texture: Dry Crystals (Y/N)                      | <u>Y</u> |
| 3. Yellowish color when dissolved in methanol (Y/N) | <u>Y</u> |
| 4. Carbamates peak areas meet specification (Y/N)   | <u>Y</u> |
| 5. Baseline noise (0 +/- 25 mvolts)/(Y/N)           | <u>Y</u> |

**Sample:** Carbamate Test Mixture, PN:1700-0063

**Storage Conditions:** Dry at 8°C to 25°C

**Analysis Certified by:** David Mazawa      **Date:** 26 April 2024  
**Title:** QC Manager