



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Amino Acids Analysis

OD104

Product: OPA Diluent for Post-column Amino Acids analysis

Part Number: OD104

Lot Number: 010052

Production Date: October 26, 2020

Expiration Date: October 26, 2022

Primary Application: Post-column analysis of amino acids.

Acceptance Criteria:

- | | |
|--|----------|
| 1. pH = 10.5 ± 0.1 @ 25 ± 0.4 °C (Y/N) | <u>Y</u> |
| 2. Peaks visible and eluted in order (Y/N) | <u>Y</u> |
| 3. Baseline noise (0 ± 25 mvolts) (Y/N) | <u>Y</u> |

Sample: Glyphosate Test Mixture, **PN** 1700-0080,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** 11/11/2020
President
Title