



CERTIFICATE OF ANALYSIS

OPA Diluent for Post-column Amino Acids Analysis

OD104

Product: OPA Diluent for Post-column Amino Acids analysis

Part Number: OD104

Lot Number: 108035

Production Date: 31 Aug 2011

Expiration Date: 30 Aug 2013

Primary Application: Post-column analysis of amino acids.

Acceptance Criteria:

1. pH = 10.5 ± 0.1 @ 25 ± 0.4 °C (Y/N)

Y

2. Peaks visible and eluted in order (Y/N)

Y

4. Baseline noise (0 ± 25 mvolts) (Y/N)

Y

Sample: Glyphosate Test Mixture, PN 1700-0080,

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

[Signature]
VP Opr

Date:

20 Sep 2011

Title
ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE