



CERTIFICATE OF ANALYSIS

GTX and STX Mobile Phase A

PSP-0001

Product: GTX and STX Mobile Phase A

Part Number: PSP-0001

Lot Number: 2510011

Production Date: 06 Oct 2025

Expiration Date: 06 Oct 2026

Primary Application: Mobile Phase for Paralytic Shellfish Toxins Analysis

Acceptance Criteria:

- | | |
|--|----------|
| 1. pH = 7.10 ± 0.05 @ 25 ± 0.4 °C (Y/N) | <u>Y</u> |
| 2. Separation of GTX and STX toxins meets specifications (Y/N) | <u>Y</u> |

Storage Conditions: Dry at 8 to 25°C

Analysis Certified by:	<u>David Mazawa</u>	Date:	<u>13 Oct 2025</u>
Title:	<u>QC Manager</u>		