



CERTIFICATE OF ANALYSIS

Post-Column Oxidant Reagent (2-part)

PSP-R1

Product: Post-Column Oxidant Reagent (2-part)

Part Number: PSP-R1

Lot Number: 2510013

Production Date: 08 Oct 2025

Expiration Date: 08 Oct 2026

Primary Application: Post-Column Reagent for Paralytic Shellfish Toxins Analysis

Acceptance Criteria:

- | | |
|--|----------|
| 1. Prepared Reagent pH = 7.8 ± 0.05 @ 25 ± 0.4 °C (Y/N) | <u>Y</u> |
| 2. All appropriate peaks on the chromatogram of GTX/STX standard are visible (Y/N) | <u>Y</u> |

Storage Conditions: Dry at 8 to 25°C

Analysis Certified by:	<u>David Mazawa</u>	Date:	<u>13 Oct 2025</u>
Title:	<u>QC Manager</u>		