



CERTIFICATE OF ANALYSIS

Regenerant for Lithium System

RG003

Product: RG003 Regenerant for Lithium System

Part Number: RG003

Lot Number: 108006

Test Start Date: 05 Aug 2011

Expiration Date: 05 Aug 2013

Primary Application: Regenerant for Lithium System.

Acceptance Criteria:

1. pH = 12.60 ± 0.20 @ 25 ± 0.4 °C
2. Conductivity 40.4 ± 0.5 (Y/N)

Actual Values

4
4

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

[Signature]
VP Oper

Title

Date: _____

20 Sep 2011

Rev. 7, 8/11/08

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE