



CERTIFICATE OF ANALYSIS

Column Regenerant for Lithium-based Amino Acids Analysis

RG003

Product: Lithium Column Regenerant

Part Number: RG003

Lot Number: 2103046

Production Date: 17 Mar 2021

Expiration Date: 17 Mar 2023

Primary Application: For column regeneration in lithium-based amino acids analysis

Acceptance Criteria:

- | | |
|--|----------|
| 1. pH = 12.60 ± 0.20 @ 25 ± 0.4 °C (Y/N) | <u>Y</u> |
| 2. Conductivity meets specifications (Y/N) | <u>Y</u> |

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** April 6, 2021
President
Title