



## CERTIFICATE OF ANALYSIS

### Column Regenerant for Lithium-based Amino Acids Analysis

**RG003**

**Product:** Lithium Column Regenerant

**Part Number:** RG003

**Lot Number:** 2210045

**Production Date:** 26 October 2022

**Expiration Date:** 26 October 2024

**Primary Application:** For column regeneration in lithium-based amino acids analysis

**Acceptance Criteria:**

1. pH =  $12.60 \pm 0.20$  @  $25 \pm 0.4$  °C (Y/N) Y

2. Conductivity meets specifications (Y/N) Y

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Maria Ofitserova **Date:** 26 October 2022  
**Title:** Senior Research Chemist