



CERTIFICATE OF ANALYSIS

Column Regenerant for Lithium-based Amino Acids Analysis

RG003

Product: Lithium Column Regenerant

Part Number: RG003

Lot Number: 2401045

Production Date: 02 Feb 2024

Expiration Date: 02 Feb 2026

Primary Application: For column regeneration in lithium-based amino acids analysis

Acceptance Criteria:

1. pH = 12.60 ± 0.20 @ 25 ± 0.4 °C (Y/N) Y

2. Conductivity meets specifications (Y/N) Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: David Mazawa **Date:** 13 Feb 2024
Title: Technical Support Chemist