



CERTIFICATE OF ANALYSIS

Column Regenerant for Lithium-based Amino Acids Analysis

RG003

Product: Lithium Column Regenerant

Part Number: RG003

Lot Number: 801007

Production Date: 08 Jan 2018

Expiration Date: 08 Jan 2020

Primary Application: For column regeneration in lithium-based amino acids analysis

Acceptance Criteria:

1. pH = 12.60 ± 0.20 @ 25 ± 0.4 °C (Y/N) y

2. Conductivity meets specifications (Y/N) y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** January 9, 2018
President
Title