



## CERTIFICATE OF ANALYSIS

### Column Regenerant for Lithium-based Amino Acids Analysis

**RG003**

**Product:** Lithium Column Regenerant

**Part Number:** RG003

**Lot Number:** 904006

**Production Date:** 02 Apr 2019

**Expiration Date:** 02 Apr 2021

**Primary Application:** For column regeneration in lithium-based amino acids analysis

**Acceptance Criteria:**

1. pH =  $12.60 \pm 0.20$  @  $25 \pm 0.4$  °C (Y/N) y

2. Conductivity meets specifications (Y/N) y

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** April 10, 2019  
President  
Title