



# CERTIFICATE OF ANALYSIS

## Regenerant for Sodium System RG011 / 1700-1120

Product: Regenerant for Sodium System

Part Number: RG011 / 1700-1120

Lot Number: 006030

Test Date: 24 Jun 2010

Expiration Date: 24 Jun 2012

Primary Application: Sodium-based amino acid analysis

### Acceptance Criteria:

### Actual Values:

1. pH =  $12.74 \pm 0.10$  @  $25 \pm 0.4$  °C
2. Conductivity =  $31.3 \pm 0.7$  @  $25 \pm 0.4$  °C

12.672  
31.1

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: \_\_\_\_\_

*Jim Murphy*  
VP Operations  
Title

Date: 08 Jul 2010

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 9, 7/09/08