



## CERTIFICATE OF ANALYSIS

### Regenerant for Sodium System RG011 / 1700-1120

**Product:** Regenerant for Sodium System

**Part Number:** RG011 / 1700-1120

**Lot Number:** 108029

**Production Date:** 24 Aug 2011

**Expiration Date:** 26 Aug 2013

**Primary Application:** Sodium-based amino acid analysis

#### Acceptance Criteria:

1. pH =  $12.74 \pm 0.10$  @  $25 \pm 0.4$  °C (Y/N)
2. Conductivity =  $31.3 \pm 0.7$  @  $25 \pm 0.4$  °C (Y/N)

4  
4

**Storage Conditions:** Dry at 8 to 25 °C

Analysis Certified by: \_\_\_\_\_

*V.P. DOR*  
\_\_\_\_\_  
Title

Date: \_\_\_\_\_

21 Sep 2011

## ACCEPTED

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev.10, 04/25/2011