



CERTIFICATE OF ANALYSIS

Column Regenerant for Ion-exchange chromatography of Sodium based Amino Acids Analysis RG011 / 1700-1120

Product: Column Regenerant for Sodium Based Analysis
Part Number: RG011 / 1700-1120
Lot Number: 110033
Production Date: 31 Oct 2011
Expiration Date: 31 Oct 2013

Primary Application: Sodium-based amino acid analysis

Acceptance Criteria:

1. pH = 12.74 ± 0.10 @ 25 ± 0.4 °C (Y/N)
2. Conductivity = 31.3 ± 0.7 @ 25 ± 0.4 °C (Y/N)

4
14

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim Murphy
VP Oper
Title

Date: _____

08 Nov 2011

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev.10, 04/25/2011