



## CERTIFICATE OF ANALYSIS

### Column Regenerant for Sodium-based Amino Acids Analysis

**RG011**

**Product:** Sodium Column Regenerant

**Part Number:** RG011

**Lot Number:** 202016

**Production Date:** 13 Feb 2012

**Expiration Date:** 13 Feb 2014

**Primary Application:** For column regeneration in sodium-based amino acids analysis

#### Acceptance Criteria:

1. pH =  $12.74 \pm 0.10$  @  $25 \pm 0.4$  °C (Y/N)
2. Conductivity meets specifications (Y/N)

4  
4

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:**

*[Signature]*  
VP Lab  
Title

**Date:**

28 Feb 2012

**ACCEPTED**

Rev. 11, 11/10/2011

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

P:\Documentation\Chemical Documentation\RG011