



## CERTIFICATE OF ANALYSIS

### Column Regenerant for Sodium-based Amino Acids Analysis

**RG011**

**Product:** Sodium Column Regenerant

**Part Number:** RG011

**Lot Number:** 2306025

**Production Date:** 15 June 2023

**Expiration Date:** 15 June 2025

**Primary Application:** For column regeneration in sodium-based amino acids analysis

**Acceptance Criteria:**

- |  |          |
|--|----------|
| 1. pH = $12.74 \pm 0.10$ @ $25 \pm 0.4$ °C (Y/N) | <u>Y</u> |
| 2. Conductivity meets specifications (Y/N)       | <u>Y</u> |

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** David Mazawa      **Date:** 19 June 2023  
**Title:** Technical Support Chemist