



CERTIFICATE OF ANALYSIS

Column Regenerant for Sodium-based Amino Acids Analysis

RG011

Product: Sodium Column Regenerant

Part Number: RG011

Lot Number: 2510057

Production Date: 30 Oct 2025

Expiration Date: 30 Oct 2027

Primary Application: For column regeneration in sodium-based amino acids analysis

Acceptance Criteria:

1. pH = 12.74 ± 0.10 @ 25 ± 0.4 °C (Y/N)

Y

2. Conductivity meets specifications (Y/N)

Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: David Mazawa
Title: QC Manager

Date: 05 Nov 2025