



CERTIFICATE OF ANALYSIS

Column Regenerant for Sodium-based Amino Acids Analysis

RG011

Product: Sodium Column Regenerant

Part Number: RG011

Lot Number: 2603027

Production Date: 13 Mar 2026

Expiration Date: 13 Mar 2027

Primary Application: For column regeneration in sodium-based amino acids analysis

Acceptance Criteria:

- | | |
|--|----------|
| 1. pH = 12.74 ± 0.10 @ 25 ± 0.4 °C (Y/N) | <u>Y</u> |
| 2. Conductivity meets specifications (Y/N) | <u>Y</u> |

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: David Mazawa **Date:** 16 Mar 2026
Title: QC Manager