



## CERTIFICATE OF ANALYSIS

### Column Regenerant for Sodium-based Amino Acids Analysis

RG011

**Product:** Sodium Column Regenerant

**Part Number:** RG011

**Lot Number:** 410001

**Production Date:** 01 Oct 2014

**Expiration Date:** 01 Oct 2016

**Primary Application:** For column regeneration in sodium-based amino acids analysis

**Acceptance Criteria:**

1. pH =  $12.74 \pm 0.10$  @  $25 \pm 0.4$  °C (Y/N)

y

2. Conductivity meets specifications (Y/N)

y

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy

President

Title

**Date:** 03 Oct 2014