



CERTIFICATE OF ANALYSIS

Column Regenerant for Sodium-based Amino Acids Analysis

RG011

Product: Sodium Column Regenerant

Part Number: RG011

Lot Number: 511005

Production Date: 05 Nov 2015

Expiration Date: 05 Nov 2017

Primary Application: For column regeneration in sodium-based amino acids analysis

Acceptance Criteria:

- 1. pH = 12.74 ± 0.10 @ 25 ± 0.4 °C (Y/N) y
- 2. Conductivity meets specifications (Y/N) y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy **Date:** November 11,2015
President
Title