



CERTIFICATE OF ANALYSIS

SERAPREP

SP100 (1800-0020)

Product: SERAPREP
Part Number: SP100 (1800-0020)
Lot Number: 012037
Test Start Date: 05 JAN 2011

Primary Application: Prepare serum and other native samples for amino acid analysis.

Acceptance Criteria:

1. pH = 1.04 ± 0.02 @ 25 °C

Actual Values

1.030

2. Use threonine and serine peaks for evaluation. Serum sample peak times ± 30 seconds of calibration standard. (Y/N)

Y

Calibration Standard:

Physiological Fluid Calibration Standard, PN 011006p, LN 909011

Analysis Certified by:

[Signature]
VP QPR

Date:

11 Feb 2011

Title

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE