



CERTIFICATE OF ANALYSIS

SERAPREP

SP100

Product: SERAPREP
Part Number: SP100 (1800-0020)
Lot Number: 110028
Production Date: 31 Oct 2011
Expiration Date: 28 Oct 2012

Primary Application: Prepare serum and other native samples for amino acid analysis.

Acceptance Criteria:

1. pH = 1.04 ± 0.02 @ 25 °C (Y/N) y

2. Use threonine and serine peaks for evaluation.
Serum sample peak times ± 30 seconds
of calibration standard. (Y/N) y

Calibration Standard:

Physiological Fluid Calibration Standard, PN: 011006P

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy

Date: 31 Oct 2011

VP OPR
Title

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE