



CERTIFICATE OF ANALYSIS

SERAPREP

SP100

Product: SERAPREP

Part Number: SP100

Lot Number: 6212 203007

Production Date: 06 MAR 2012

Expiration Date: 06 MAR 2013

Primary Application: Preparation of serum and other native samples for amino acids analysis

Acceptance Criteria:

1. pH is 1.04 ± 0.02 at $25^\circ \text{C} \pm 0.4^\circ \text{C}$ (Y/N)
2. Retention times difference for Threonine and Serine meets specification (Y/N)
3. Baseline: no peaks (Y/N)

Y
Y
Y

Sample: Serum sample prepared with SERAPREP

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:

[Signature]
VP Ops
Title

Date:

19 MAR 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE