



# CERTIFICATE OF ANALYSIS

## SERAPREP

### SP100

**Product:** SERAPREP

**Part Number:** SP100

**Lot Number:** 209034

**Production Date:** 26 Sep 2012

**Expiration Date:** 26 Sep 2013

**Primary Application:** Preparation of serum and other native samples for amino acids analysis

#### Acceptance Criteria:

1. pH is  $1.04 \pm 0.02$  at  $25^{\circ}\text{C} \pm 0.4^{\circ}\text{C}$  (Y/N)

y

2. Retention times difference for Threonine and Serine meets specification (Y/N)

y

3. Baseline: no peaks (Y/N)

y

**Sample:** Serum sample prepared with SERAPREP

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy

**Date:** 05 Oct 2012

VP Operations

Title