

CERTIFICATE OF ANALYSIS SERAPREP

SP100

Product: SERAPREP			
Part Number: SP100 Lot Number:	2209050		
Production Date:	26 Sep 2022		
Expiration Date:	26 Sep 2023		
Primary Application: Preparation of serum and other native samples for amino acids analysis			
Acceptance Criteria:			
1. pH is 1.04 ± 0.02 at 25° C ± 0.4°C (Y/N)			Υ
•	rence for Threonine and Serine	-	-
meets specification	(Y/N)	_	Υ
3. Baseline: no peaks (Y/N)	_	Υ
Sample: Serum sample prepared with SERAPREP			
Storage Conditions: Dry at 8 to 25 °C			
Analysis Certified by:	David Mazawa	Date:	28 Sep 2022
Title:	Technical Support Chemist		