



## CERTIFICATE OF ANALYSIS

**SERAPREP**

**SP100**

**Product:** SERAPREP

**Part Number:** SP100

**Lot Number:** 2511027

**Production Date:** 17 Nov 2025

**Expiration Date:** 17 Nov 2026

**Primary Application:** Preparation of serum and other native samples for amino acids analysis

**Acceptance Criteria:**

1. pH is $1.04 \pm 0.02$ at $25^{\circ} \text{C} \pm 0.4^{\circ} \text{C}$ (Y/N)	<u>Y</u>
2. Retention times difference for Threonine and Serine meets specification (Y/N)	<u>Y</u>
3. Baseline: no peaks (Y/N)	<u>Y</u>

**Sample:** Serum sample prepared with SERAPREP

**Storage Conditions:** Dry at 8 to 25 °C

<b>Analysis Certified by:</b>	<u>David Mazawa</u>	<b>Date:</b>	<u>25 Nov 2025</u>
<b>Title:</b>	<u>QC Manager</u>		