



CERTIFICATE OF ANALYSIS

SERAPREP

SP100

Product: SERAPREP

Part Number: SP100

Lot Number: 308008

Production Date: 07 Aug 2013

Expiration Date: 07 Aug 2014

Primary Application: Preparation of serum and other native samples for amino acids analysis

Acceptance Criteria:

1. pH is 1.04 ± 0.02 at $25^{\circ}\text{C} \pm 0.4^{\circ}\text{C}$ (Y/N)

y

2. Retention times difference for Threonine and Serine meets specification (Y/N)

y

3. Baseline: no peaks (Y/N)

y

Sample: Serum sample prepared with SERAPREP

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: Jim Murphy
President

Title

Date: 26 August 2013