



# CERTIFICATE OF ANALYSIS

## SERAPREP

### SP100

**Product:** SERAPREP

**Part Number:** SP100

**Lot Number:** 910010

**Production Date:** 07 Oct 2019

**Expiration Date:** 07 Oct 2020

**Primary Application:** Preparation of serum and other native samples for amino acids analysis

#### Acceptance Criteria:

- |  |          |
|--|----------|
| 1. pH is $1.04 \pm 0.02$ at $25^{\circ} \text{C} \pm 0.4^{\circ} \text{C}$ (Y/N) | <u>y</u> |
| 2. Retention times difference for Threonine and Serine meets specification (Y/N) | <u>y</u> |
| 3. Baseline: no peaks (Y/N)  | <u>y</u> |

**Sample:** Serum sample prepared with SERAPREP

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** Jim Murphy **Date:** October 15, 2019  
President  
Title