



# CERTIFICATE OF ANALYSIS


## TRIONE, Ninhydrin Reagent T100c

Product: TRIONE, Ninhydrin Reagent  
Part Number: T100c  
Lot Number: 007009  
Test Start Date: 12 Jul 2010  
Expiration Date: 08 Oct 2010

Primary Application: Amino acid analysis

<b>Acceptance Criteria:</b> 1. <b>TRIONE</b> color without shades of brown or purple (Y/N) 2. L-aspartic acid to L-proline ratio $\geq 3.3$	<b>Actual Values:</b> <u>4</u> <u>4</u>
---	---

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by:  Date: 15 Jul 2010  
VP OPR  
 Title

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 7, 8/12/08