



## CERTIFICATE OF ANALYSIS

### TRIONE, Ninhydrin Reagent T100/T100C

**Product:** TRIONE, Ninhydrin Reagent

**Part Number:** T100/T100C

**Lot Number:** 112021

**Production Date:** 19 Dec 2011

**Expiration Date:** 15 MAR 2012

**Primary Application:** Amino acid analysis


**Acceptance Criteria:**

1. TRIONE color without shades of brown or purple (Y/N)
2. L-aspartic acid to L-proline ratio  $\geq 3.0$  (Y/N)

Y  
Y

**Storage Conditions:** Dry at 8 to 25 °C

**Sample:** Aspartic acid and Proline mixture.

**Analysis Certified by:**   
VP  
Title

**Date:** 28 Feb 2012

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 8, 09/15/11

P:\Documentation\Chemical Documentation\T100C