



CERTIFICATE OF ANALYSIS

TRIONE Ninhydrin Reagent T100/T100C

Product: TRIONE, Ninhydrin Reagent

Part Number: T100/T100C

Lot Number:

20 202B

Production Date:

2/23/2012

Expiration Date:

5/21/2012

Primary Application: Amino acid analysis

Acceptance Criteria:

1. TRIONE color without shades of brown or purple (Y/N)
2. L-aspartic acid to L-proline ratio ≥ 3.0 (Y/N)

Y
Y

Storage Conditions: Dry at 8 to 25 °C

Sample: Aspartic acid and Proline mixture.

Analysis Certified by:

[Signature]
VP OPR
Title

Date:

05 MAR 2012

ACCEPTED

PICKERING LABORATORIES, INC.
QUALITY ASSURANCE

Rev. 8, 09/15/11

P:\Documentation\Chemical Documentation\F100C