



## CERTIFICATE OF ANALYSIS

### TRIONE, Ninhydrin Reagent (2 Parts) T200

**Product:** TRIONE, Ninhydrin Reagent (2 Parts)

**Part Number:** T200

**Lot Number:** 005006

**Test Start Date:** 09 MAY 2010

**Expiration Date:** 05 MAY 2011

**Primary Application:** Sodium amino acid analysis

**Acceptance Criteria:**

1. TRIONE color without shades of brown or purple.
2. L-aspartic acid to L-proline ratio  $\geq 3.3$

**Actual Values:**

4  
4

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** \_\_\_\_\_

*[Signature]*  
VP OPR  
\_\_\_\_\_  
Title

**Date:** 10 MAY 2010

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 6, 8/11/08