



# CERTIFICATE OF ANALYSIS

## TRIONE, Ninhydrin Reagent (2 Parts) T200

Product: TRIONE, Ninhydrin Reagent (2 Parts)

Part Number: T200

Lot Number: 008030

Test Start Date: 24 Aug 2010

Expiration Date: 20 Aug 2011

Primary Application: Sodium amino acid analysis

### Acceptance Criteria:

1. TRIONE color without shades of brown or purple.
2. L-aspartic acid to L-proline ratio  $\geq 3.3$

### Actual Values:

4  
4

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: \_\_\_\_\_

*Jim Murphy*  
VP QPC

Title

Date: \_\_\_\_\_

30 Aug 2010

Rev. 6, 8/11/08

**ACCEPTED**  
PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE