



CERTIFICATE OF ANALYSIS

TRIONE, Ninhydrin Reagent (2 Parts) T200

Product: TRIONE, Ninhydrin Reagent (2 Parts)

Part Number: T200

Lot Number: 012033

Test Start Date: 16 DEC 2010

Expiration Date: 15 DEC 2011

Primary Application: Sodium amino acid analysis

Acceptance Criteria:

1. TRIONE color without shades of brown or purple.
2. L-aspartic acid to L-proline ratio ≥ 3.3

Actual Values:

4
4

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: *Jim Murphy* Date: 18 JAN 2011

VP OPR
Title

Rev. 6, 8/11/08

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE