



CERTIFICATE OF ANALYSIS

**TRIONE, Ninhydrin Reagent (2 Parts)
T200**

Product: TRIONE, Ninhydrin Reagent (2 Parts)

Part Number: T200

Lot Number: 103019

Production Date: 11 MAR 2011

Expiration Date: 10 MAR 2012

Primary Application: Sodium amino acid analysis

Acceptance Criteria:

1. **TRIONE** color without shades of brown or purple(Y/N)
2. L-aspartic acid to L-proline ratio ≥ 3.3 (Y/N)

Y
Y

Storage Conditions: Dry at 8 to 25 °C

Analysis Certified by: _____

Jim Murphy
VP OPR
Title

Date: 23 MAR 2011

Rev. 7, 3/25/2011

ACCEPTED
PICKERING LABORATORIES, INC.
QUALITY ASSURANCE