



**CERTIFICATE OF ANALYSIS**  
**TRIONE, Ninhydrin Reagent (2 Parts)**  
**T200**

**Product:** TRIONE, Ninhydrin Reagent (2 Parts)

**Part Number:** T200

**Lot Number:** 103049

**Production Date:** 01 APR 2011

**Expiration Date:** 01 APR 2012

**Primary Application:** Sodium amino acid analysis

**Acceptance Criteria:**

1. **TRIONE** color without shades of brown or purple(Y/N)
2. L-aspartic acid to L-proline ratio  $\geq 3.3$  (Y/N)

4  
4

**Storage Conditions:** Dry at 8 to 25 °C

**Analysis Certified by:** \_\_\_\_\_

*Jim Myrb*  
VP OPER

Title

**Date:** \_\_\_\_\_

18 APR 2011

**ACCEPTED**

PICKERING LABORATORIES, INC.  
QUALITY ASSURANCE

Rev. 7, 3/25/2011